

Amperometric and spectrophotometric determination of carbaryl in natural waters and commercial formulations

Dionísia C. Portela · Isabel M. F. Pereira · Paula Paíga Cristina Delerue-Matos · M. Carmo V. F. Vaz

Abstract The work presented describes the development and evaluation of two flow-injection analysis (FIA) systems for the automated determination of carbaryl in spiked natural waters and commercial formulations. Samples are injected directly into the system where they are subjected to alkaline hydrolysis thus forming 1-naphthol. This product is readily oxidised at a glassy carbon electrode. The electrochemical behaviour of 1-naphthol allows the development of an FIA system with an amperometric detector in which 1-naphthol determination, and thus measurement of carbaryl concentration, can be performed. Linear response over the range 1.0×10^{-7} to 1.0×10^{-5} mol L⁻¹, with a sampling rate of 80 samples h⁻¹, was recorded. The detection limit was 1.0×10^{-8} mol L⁻¹. Another FIA manifold was constructed but this used a colorimetric detector. The methodology was based on the coupling of 1-naphthol with phenylhydrazine hydrochloride to produce a red complex which has maximum absorbance at 495 nm. The response was linear from 1.0×10^{-5} to 1.5×10^{-3} mol L⁻¹ with a detection limit of 1.0×10^{-6} mol L⁻¹. Sample-throughput was about 60 samples h⁻¹. Validation of the results provided by the two FIA methodologies was performed by comparing them with results from a standard HPLC–UV technique. The relative deviation was <5%. Recovery trials were also carried out and the values obtained ranged from 97.0 to 102.0% for both methods. The repeatability (*RSD*, %) of 12 consecutive injections of one sample was 0.8% and 1.6% for the amperometric and colorimetric systems, respectively.

Keywords Amperometry · Spectrophotometry · Carbaryl · Pesticides · Flow-injection analysis

Introduction

The carbamates belong to a group of pesticides that has attained great popularity in recent years due to their broad biological activity [1]. They are used as insecticides, miticides, fungicides and molluscicides [1]. Because of the wide range of uses in the treatment of seeds, soils, or crops [1], the methyl carbamates constitute one of the most important classes in this group and of these, carbaryl (1-naphthyl *N*-methyl carbamate) has been one of the most used because it has low oral and skin toxicity in spite of its great insecticide capacity [2].

Most of the analytical methods employed for the quantitation of carbaryl have been based on chromatographic techniques. Classical gas chromatography has been shown to be generally unsatisfactory due to the thermal instability of carbaryl, requiring either a chemical derivation step (use of different reagents) or the employment of short columns, specially treated columns or short capillaries [1]. Thus many authors have preferred to make use of liquid chromatography (generally reversed-phase) linked to various detectors, e.g. UV [3], diode-array [4, 5], fluorescence

[5, 6], or electrochemical [7], although some of these methods also require pre-treatment steps to form detectable derivatives. Thin-layer chromatography has occasionally been employed [8].

Other methods for estimation of carbaryl in various matrices are found in the literature. These include: spectrophotometric techniques using UV/vis [9, 10, 11, 12],

infrared spectrometry [13], and electrochemistry [1, 2]. All these are founded on the conversion of carbaryl to 1-naphthol by means of alkaline hydrolysis. The spectrophotometric methods using UV/vis detection mention several chromogenic reagents that form coloured complexes with

1-naphthol in order to achieve an appropriate selectivity and sensitivity of the spectrophotometric measurements. The ease with which 1-naphthol is oxidised allows satisfactory detection using electrochemistry based on a differential pulse voltammetric method [1]. The coulometric oxidation product of carbaryl at a platinum electrode (1,4-naphthoquinine) can also be determined electrochemically.

This is performed either directly by adsorptive stripping voltammetry or indirectly by differential pulse polarography after reduction of the oxidation product (1,4-naphtho-quinine) at a dropping mercury electrode [2].

The increasingly widespread use of pesticides, frequently unregulated, linked to the growing concern for environmental problems has made the determination of the concentrations of these compounds progressively more frequent and rigorous. To respond to these demands control laboratories have had to introduce automated analyses that reduce substantially the time required for analysis and allow the development of high-throughput capability.

Flow-injection analysis (FIA) can provide this requirement while presenting as an additional advantage the possibility of employing common laboratory equipment and reducing drastically reagent consumption as compared with traditional batch procedures. FIA coupled to a spectrophotometric detector has already been used by Boaventura et al. [14] and Ricardo et al. [15]. In the first article 1-naphthol was reacted with *p*-aminophenol to yield a blue complex with maximum absorbance at 596 nm. The methodology developed by Ricardo et al. [15], is based on the preconcentration of carbaryl into a polyether type polyurethane foam followed by on-line elution with dichloromethane and measurement of absorbance at 280 nm. An amperometric flow-injection biosensor device for determination of carbaryl by means of 4-aminophenol quantification was also proposed. This xenobiotic agent was here responsible for the inhibition of the acetylcholinesterase activity toward the substrate 4-aminophenylacetate. The enzyme was immobilised at the surface of a glassy carbon electrode thus recording the decrease of the amount of 4-aminophenol [16].

The present work reports on the development of two FIA systems which enable determinations of carbaryl in two different ranges of concentration without prior treatment of samples. Alkaline hydrolysis of carbaryl to 1-naphthol is carried out inside both systems and the latter compound is quantitated. The first system makes use of the ease of electrochemical oxidation of 1-naphthol at a glassy carbon electrode and has an amperometric detector. In the other system 1-naphthol reacts with phenylhydrazine hydrochloride to give a red complex with absorbance measurement at 495 nm.

solutions of 1-naphthol were prepared by diluting the stock solution with water.

Acetate buffer solutions of pH from 4.0 to 5.9 were used as support electrolyte and were prepared by mixing different volumes of acetic acid and sodium acetate solutions, both 2.0 mol L^{-1} , until the desired pH was reached. Subsequent dilution was performed to furnish solutions with a final ionic strength of 0.2 mol L^{-1} .

In the comparison method (chromatographic determination) solvents were of HPLC grade. Before use they were filtered and any dissolved air was removed by bubbling helium through the solution.

Standard and sample preparation

Stock solutions of carbaryl ($1.0 \times 10^{-2} \text{ mol L}^{-1}$) were prepared with an exact weight of the pure pesticide, dissolved in the least quantity of *N,N*-dimethylformamide (DMF) and diluting to 10.00 mL with water. The resulting solution was kept in the dark at $+4^\circ\text{C}$. The standard solutions used for the optimisation studies and plotting calibration curves were prepared by dilution of these stock solutions with water. These solutions were stable for at least one week if kept in the dark at $+4^\circ\text{C}$ when not in use.

Natural water samples were collected from various locations in Porto (rivers and lakes) in dark glass bottles. The samples were spiked with carbaryl and directly analysed by the different methodologies using the calibration curve method.

In the commercial product (Permutex, Permutadora) the proportion of carbaryl present is 50% (w/w); therefore the sample stock solution ($3.0 \times 10^{-4} \text{ mol L}^{-1}$) was prepared by accurately weighing 0.0060 g of the commercial sample and adding a few drops of DMF until complete dilution. Water was then added and the total volume was adjusted to 50.00 mL. The sample stock solution was diluted with water in order to obtain a concentration within the calibration curve range. In the spectrophotometric system the formulations were passed through a $0.45 \text{ }\mu\text{m}$ membrane filter before introduction into the system, to avoid turbidity or particulate contamination that might affect the analytical signal.

Apparatus

HPLC readings under batch conditions were made by means of a Sykam A 1210 liquid chromatograph equipped with a model 3200 UV detector tuned to 222 nm. Separation of sample components was accomplished on a Supercosil LC-18 column (250 mm \times 4.6 mm, 5 μm particle size) from Macherey-Nagel, Germany.

The FIA systems comprised a Gilson Minipuls 3 peristaltic pump, fitted with PVC tubing (1.0 mm i.d.) and a four-way Rheodyne type 5020 injection valve. PTFE tubing (Omnifit, Teflon, 0.8 mm i.d.) and Gilson end-fittings and connectors were used to connect all the components of the manifold. Other auxiliary devices, such as Perspex Y-shaped confluences constructed as previously described [17], were also used.

The output signals were recorded on a Kipp and Zonen BD 112 recorder.

All other chemicals were Merck pro analysis grade and all solutions were prepared using purified water (conductivity $<0.1 \text{ }\mu\text{S cm}^{-1}$) obtained from a Barnstead E-pure 4 system.

The phenylhydrazine hydrochloride solution (0.3% w/v) was prepared from the corresponding solid and dissolved in water. This solution was prepared daily and

Experimental

Reagents and solutions

Carbaryl (Pestanal grade, 99.9%) was purchased from Riedel-de Haën and used without further purification.

kept in a closed container away from light in order to avoid oxidation.


Stock solutions of 1-naphthol (1.0×10^{-2} mol L⁻¹) were prepared by careful weighing of the solid, dissolution in the least quantity of ethanol, and diluting to volume with water. More dilute working

The spectrophotometric detector was a Jenway 6300 single-beam unit comprising a 100- μ L flow-through cell and an optical path length of 10 mm.

Amperometric detection was performed by use of a 641 VA Metrohm detector linked to a 656 Metrohm wall-jet containing a three-electrode system – a glassy carbon working electrode (Metrohm 6.0805.010) ($d=3.0$ mm), an Ag/AgCl/KCl 3.00 mol L⁻¹ reference electrode (Metrohm 6.07027.000), and a gold counter-electrode (Metrohm 6.530.320).

When necessary the working electrode was mechanically cleaned by polishing its surface using the specified polishing kit (Metrohm 6.2802.010), first with γ -Al₂O₃ (0.3 μ m) until a shining surface was obtained and afterwards only with water.

The pH of buffer solutions was determined with a Crison 2002 pH meter with a Sentek 71728 combined glass electrode.



volume (215 μ L); R_1 , carrier stream (water); R_2 , sodium hydroxide solution (2.0 mol L^{-1}); R_3 , phenylhydrazine solution 0.3%; X , Y , confluence points; L_n , reactors ($L_1=30 \text{ cm}$, $L_2=700 \text{ cm}$); $q_2=q_3=0.7 \text{ mL min}^{-1}$, $q_1=1.4 \text{ mL min}^{-1}$; Rec , recorder; W , waste

Fig. 1 Flow-injection system with amperometric detection: P , peristaltic pump; S , sample; I , injection volume (215 μ L); R_1 , carrier stream (water); R_2 , sodium hydroxide solution (0.02 mol L^{-1}); R_3 , acetate buffer solution ($\text{pH}=5.0$); X , Y , confluence points; L_n , reactors ($L_1=30 \text{ cm}$, $L_2=250 \text{ cm}$); $q_1=q_2=1.0 \text{ mL min}^{-1}$, $q_3=2.0 \text{ mL min}^{-1}$; DET , detector; Rec , recorder; W , waste

Comparison method

Results from amperometric and spectrophotometric analysis were compared with those obtained using an independent method employed by Riedel-de Haën for quality control of pro analysis grade reagent [18]. HPLC was performed at room temperature with a mixture of water (60%) and acetonitrile (40%) as mobile phase at a flow rate of 1.35 mL min^{-1} . Calibration was performed by injection of 20 μ L of carbaryl standard solutions, with concentrations of 1.0×10^{-6} to $8.0 \times 10^{-6} \text{ mol L}^{-1}$ and 5.0×10^{-5} to $5.0 \times 10^{-4} \text{ mol L}^{-1}$ for the amperometric and spectrophotometric systems, respectively.

Flow-injection configuration

Amperometric detector

The manifold used for the determination of carbaryl with amperometric detection has two confluence points and is depicted in Fig. 1. A sample (I) was introduced in an ultra-pure water carrier stream (R_1) without previous treatment and the sample plug was then conveyed to the confluence X where NaOH solution (R_2) is added to the flow. Alkaline hydrolysis occurs in the coiled tube reactor (L_1); the bolus travels to the confluence Y where it merges with buffer acetate solution (R_3) and in reactor L_2 the ionic strength and pH adjustment are made as is required by the detector.




Fig. 2 Flow-injection system with spectrophotometric detection: P , peristaltic pump; S , sample; I , injection

Spectrophotometric detector

A schematic representation of the FIA manifold with spectrophotometric detector is shown in Fig. 2.

Each sample aliquot was injected in a water carrier stream (R_1) and merged at the X confluence with a solution of phenylhydrazine hydrochloride (R_3) and NaOH (R_2) prepared in reactor (L_1). Alkaline hydrolysis of carbaryl occurred in the coiled tube reactor (L_2) forming 1-naphthol which in turn reacts with the phenylhydrazine hydrochloride to give a red complex with absorbance measurement at 495 nm.

Results and discussion

Optimisation of the FIA systems

FIA manifolds were designed and then optimized to obtain the lowest detection limit possible while maximising sample throughput, sensitivity, and precision without sample pre-treatment. All necessary modifications to the samples were effected within the FIA system. After preliminary experiments to establish the manifold designs and parameters, a univariant optimization procedure was applied and this will be separately described.

Amperometric system

Because carbaryl is not electroactive and the hydrolysis product (1-naphthol) is readily oxidised at the glassy carbon electrode, optimization was started by investigating the effect of pH of the support electrolyte and of the working electrode potential in a simpler, double-channel system (Fig. 3) in which a standard solution of 1-naphthol (concentration $1.0 \times 10^{-5} \text{ mol L}^{-1}$) was inserted into a flow of water (R_1). At confluence X the acetate buffer solution was added (R_2) to change the pH and ionic strength in reactor L_1 before arriving at the detector. Based on the previous results of the voltammetric study [1] the potential was fixed at +0.80 V and the pH was varied from 4.0 to 5.9. Within this pH range, the best results – higher absolute response and better reproducibility – were obtained at pH 5.0.




Fig. 3 Flow-injection system used for optimization: *P*, peristaltic pump; *S*, sample; *I*, injection volume (215 μL); R_1 , carrier stream (water); R_2 , acetate buffer solution; *X*, confluence point; L_1 , reactor; q_1 , q_2 ,

With acetate buffer solution of pH 5.0 as the support electrolyte, the oxidation potential of the working electrode was varied from 0.60 to 1.30 V vs. Ag/AgCl, finding an optimum at 0.80–1.0 V. Up to 0.80 V, increasing the potential led to a significant increase in peak height. Between 0.80 and 1.0 V the signal was fairly constant, above 1.0 V the analytical response decreased significantly and reproducibility suffered accordingly. This behaviour of the detection system indicated that the range 0.80 to 1.0

V would lead to the highest possible sensitivity. To reduce possible interference from compounds other than carbaryl present in real samples, the potential of 0.80 V was chosen. The length of reactor L₁ was fixed at 30 cm to give the best compromise between sensitivity and reproducibility. Following the optimization of support electrode pH and the oxidation potential, the

next step was to study the effect of changing the concentration of sodium hydroxide used in the alkaline hydrolysis. This required the construction of an FIA system with two confluences (Fig. 1) in which the standard solutions of carbaryl (concentrations 1.0×10^{-7} to 8.0×10^{-6} mol L⁻¹) were inserted into the ultra-pure water stream and carried to confluence X where the sodium hydroxide stream was added in order for the alkaline hydrolysis to occur in the reactor (L₁).

The resulting 1-naphthol-containing flow then encountered the acetate buffer at confluence Y and pH and ionic strength adjustment proceeded in reactor L₂ before finally reaching the detector. The sodium hydroxide concentration is a fundamental parameter for the optimization of this FIA system and therefore a wide range of strengths was tested (0.005 – 0.06 mol L⁻¹). Up to 0.02 mol L⁻¹ the analytical signal increased with concentration, thus this level was chosen for subsequent runs. The degree of hydrolysis was studied for each concentration of sodium hydroxide tested by comparing the analytical signals obtained from standard solutions of 1-naphthol or carbaryl with the same concentrations. Again, the highest efficiency was noted to be at the chosen concentration of sodium hydroxide –

1.2 mol L⁻¹.

1-Naphthol was formed in reactor L₁ and so five lengths (33–500 cm) were tested. The length of 250 cm was chosen, since smaller ones decreased the sensitivity and reproducibility due, presumably, to insufficient mixing of the sample with the NaOH solution and with longer ones the sensitivity and sampling rate decreased due to the dispersion effect. Both reactors (L₁ and L₂) were coiled to improve radial mixing and minimise the dispersion of the sample plug [19].

With the purpose of selecting the most suitable injection volume, several values were tested in the range of 160 to 420 μ L and it was found that the best sensitivity and reproducibility was achieved with 215 μ L. Therefore this was the volume chosen.

The optimum flow rate at the detector was found to be 4 mL min⁻¹; although no higher flow rates were tried, because they would have exceeded the maximum flow rate allowed in the wall-jet cell [20].




Fig. 4 Flow injection signals obtained with the amperometric system corresponding to injection of a set of standards of carbaryl and samples with recoveries. A. 2.0×10^{-7} mol L⁻¹; B. 5.0×10^{-7} mol L⁻¹; C. 1.0×10^{-6} mol L⁻¹; D. 2.0×10^{-6} mol L⁻¹; E. 5.0×10^{-6} mol L⁻¹; F. 8.0×10^{-6} mol L⁻¹; G, sample 2 (in Table 1); H, recovery trial of sample 2; I, sample 1 (in Table 1); J, recovery trial of sample 1

Spectrophotometric system

The 1-naphthol–phenylhydrazine reaction occurs at strongly alkaline pH [8]. To select the appropriate pH for formation of the coloured complex, phenylhydrazine solutions of concentrations between 0.25 and 1.0% were prepared in sodium hydroxide solutions of 0.5 to 2.5 mol L⁻¹ concentration (Fig. 2, reactor L₁). The colour intensity of the complex was determined by measuring the absorbance of 1.0×10^{-5} , 1.0×10^{-4} , and 1.0×10^{-3} mol L⁻¹ standard solutions of carbaryl. The maximum absorbance was achieved with a phenylhydrazine concentration of 0.3% in 2.0 mol L⁻¹ sodium hydroxide solution.

Several characteristics of the FIA system – flow rate, injection volume and reaction coil length – were also optimized.

The length of reactor L₁ was tested from 10 to 50 cm; the value 30 cm was chosen as it allowed the best results in terms of sensitivity and reproducibility. It

seems that

Table 1 Determination of carbaryl by use of FIA and HPLC

Values are means and standard deviations from five and three determinations by FIA and HPLC, respectively
^aSpiked water samples
^bCommercial formulation

Sample	Method 1			Method 2		
	FIA / Amp ($\times 10^{-6}$ mol L ⁻¹)	HPLC ($\times 10^{-6}$ mol L ⁻¹)	RE (%)	FIA / UV/vis ($\times 10^{-4}$ mol L ⁻¹)	HPLC ($\times 10^{-4}$ mol L ⁻¹)	RE (%)
1 ^a	1.90 \pm 0.08	1.95 \pm 0.06	-2.6	2.95 \pm 0.11	3.11 \pm 0.08	-4.8
2 ^a	1.95 \pm 0.07	1.92 \pm 0.05	+1.6	3.02 \pm 0.13	3.11 \pm 0.07	-2.9
3 ^a	2.02 \pm 0.05	2.09 \pm 0.06	-3.3	3.05 \pm 0.08	3.15 \pm 0.09	-3.2
4 ^b	2.19 \pm 0.06	2.11 \pm 0.06	+3.8	1.50 \pm 0.05	1.48 \pm 0.04	+1.4

mixing of both solutions occurs very quickly and that increasing the length of this reactor results in a dilution effect.

With the first reactor set to 30 cm the length of the second (L₂) was varied from 400 to 750 cm. The sensitivity increased with increasing coil length until 700 cm, after which sensitivity started decreasing, presumably as a result of the dispersion effect.

Under the conditions already selected the injection volume was varied from 100 to 500 μ L. As expected, increasing the injection volume led to an increase in sensitivity, especially for low injection volumes. Having regard for the compromise between sensitivity, reagent consumption, and sampling rate, the injection volume of 215 μ L was selected.

Flow rates before confluence Y were equal to avoid using more concentrated and therefore less stable solutions, which is particularly important for phenylhydrazine. Various flow rates were tried between 2.0 and 4.0 mL min⁻¹ and 2.8 mL min⁻¹ was selected, because lower values reduced the sampling rate and at higher values the sensitivity was significantly affected.

deviations were lower than 2.0% for both systems.

Determination of carbaryl in samples

With the conditions as specified above, linear calibration plots were obtained over the ranges 1.0×10^{-7} – 1.0×10^{-5} mol L⁻¹ ($y = 864703x + 0.1932$ $R^2 = 0.9996$) and 1.0×10^{-5} – 1.5×10^{-3} mol L⁻¹

($y = 136.97x + 0.0005$ $R^2 = 0.9992$) for determination of carbaryl with the amperometric (Fig. 4) and spectrophotometric systems respectively. The suitability of the developed FIA systems for estimation of carbaryl was assessed by analysing four samples (three samples of spiked water, and one commercial preparation) and the mean values from five determinations, with standard deviations, are shown in Table 1. Validation of the results was confirmed by comparison with the values obtained from HPLC determinations and the relative errors were always <5%. Recovery trials ranged from 97% to 102% confirming the accuracy of the developed systems. The detection limits of the methods calculated under the optimized conditions and according to IUPAC recommendations [21] were 5.0×10^{-8} mol L⁻¹ and 5.0×10^{-6} mol L⁻¹ of pesticide for the amperometric and spectrophotometric systems respectively. The precision of the FIA methods was estimated by calculating the relative standard deviation from 12 consecutive injections of one sample. The relative standard

Conclusions

The automated FIA systems developed constitute good alternatives to conventional methods. The results are highly comparable, the sampling rates are higher, and there is considerable saving of reagents. The advantages are both environmental and economic.

It should also be noted that the equipment used is easily accessible in any control laboratory and that with the two systems proposed it is possible to quantify carbaryl without pretreatment of the samples over the range 1.0×10^{-7} mol L⁻¹ to 1.5×10^{-3} mol L⁻¹. The two systems are complementary: in the amperometric detection system concentrations of carbaryl higher than 1.0×10^{-5} mol L⁻¹ cannot be measured due to signal irreproducibility caused by adsorption problems on the surface of the working electrode; with the spectrophotometric system higher concentrations are easily measured.

Acknowledgements The authors acknowledge Fundação para a Ciência e Tecnologia and Feder for financial support (POCTI/ 35287/AGR/1999).

References

1. Guiberteau A, Díaz TG, Salinas F, Ortiz JM (1995) *Anal Chim Acta* 307:219–226
2. Pérez-López JA, Zapardiek A, Bermejo E, Arauzo E, Hernández L (1994) *Fresenius J Anal Chem* 350:620–625
3. Gil-Agusti M, Alvarez-Rodriguez L, Monferrer-Pons L, Bose D, Durgbanshi A, Esteve-Romero J (2002) *Anal Lett* 35:1721–1734
4. Toscano IAS, Ribeiro ML, Santelli SE, Guardia M (2000) *Quim Nova* 23:466–471
5. Hidalgo C, Sancho JV, Roig-Navarro A, Hernandez F (1998) *Chromatographia* 47:596–600
6. Abad A, Moreno MJ, Pelegri R, Martinez MI, Saez A, Gamon M, Montoya A (1999) *J Chromatogr A* 833:3–12
7. Díaz TG, Guiberteau A, Salinas F, Ortiz JM (1996) *J Liq Chromatogr Related Technol* 19:2681–2690
8. Patil VB, Shingare MS (1993) *J Chromatogr A* 653:181–183
9. Tunceli A, Bag H, Turker AR (2001) *Fresenius J Anal Chem* 371:1134–1138
10. Demirbas A (2000) *Environ Technol* 21:351–356
11. Demirbas A (1998) *Sci Total Environ* 220:235–241
12. Alvarez Rodriguez L, Monferrer-Pons L, Esteve Romero JS, Garcia Alvarez CMC, Ramis Ramos G (1997) *Analyst* 122: 459–463
13. Daghbouche Y, Garrigues S, Guardia M (1995) *Anal Chim Acta* 314:203–212
14. Reis BF, Morales-Rubio A, Guardia M (1999) *Anal Chim Acta* 392:265–272

15. Cassela RJ, Garrigues S, Santelli RE, Guardia M (2000) *Talanta* 52:717–725
16. La Rosa C, Pariente F, Hernández L, Lorenzo E (1995) *Anal Chim Acta* 308:129–136
17. Alegret S, Alonso J, Bartroli J, Machado AASC, Lima JLFC, Paulis JM (1987) *Quim Anal* 6:278–292
18. Riedel–de Haën (1999) Certificate of analysis, carbaryl
19. Ruzicka J, Hansen EH (1988) *Flow-injection analysis*, 2nd edn. Wiley, New York
20. Metrohm (1984) *Electrochemical detection in HPLC*. Metrohm AG, CH-9100, Herisau
21. Miller JC, Miller JN (1993) *Statistics for analytical chemistry*, 3rd edn. Ellis Horwood, New York