

Automatic Quality Assessment of Smart Device Microphone Spirometry

B. Pinho^{1,2}, R. Almeida², C. Jácome², J. P. Teixeira³, R. Amaral^{2,4}, F. Lopes¹, T. Jacinto^{1,2,4},
R. Guedes², M. Pereira^{1,2}, I. Gonçalves^{1,2} and J. A. Fonseca^{1,2}

¹MEDIDA - Serviços em Medicina, EDucação, Investigação, Desenvolvimento e Avaliação, LDA, Porto, Portugal

²CINTESIS - Centro de Investigação em Tecnologias e Serviços de Saúde, MEDCIDS,

Departamento de Medicina da Comunidade Informação e Decisão em Saúde, Faculdade de Medicina,
Universidade do Porto, Porto, Portugal

³INESC TEC Campus da FEUP, Porto, Portugal

⁴Escola Superior de Saúde, Politécnico do Porto, Porto, Portugal

Keywords: Asthma, Microphone Spirometry, Spirometry Quality Assessment.

Abstract: Lung function tests are critical for diagnosis and monitoring of asthma and other respiratory diseases. Monitoring of lung function, in the absence of a healthcare professional, is very challenging but may be obtained through Smart Devices if automated quality assessment systems guarantee the proper technique during the forced expiratory manoeuvre. This paper describes the evaluation of one such system that uses the microphone of smart devices, regarding the initial effort of forced expiratory manoeuvres using the Back Extrapolated Volume. A health professional recorded microphone spirometry in 55 children (5-10 years), using a mobile game engineered for the purpose, and registered its quality. At least one acceptable manoeuvre was achieved for 96% of the children using a featured threshold. Using a stricter threshold of 5% of forced vital capacity, it was possible to ensure at least one acceptable manoeuvre for 69%. While the obtained results are comparable to findings in literature for regular spirometry in this age group, further work is required before we can determine whether the proposed algorithm is effective in real life.

1 INTRODUCTION

Spirometry is the most widely used non-invasive test of lung function, used for detection and diagnosis of various respiratory diseases, including asthma, in children (Pierce, 2005). The performance of a forced expiratory manoeuvre (FEM) involves three distinct phases: maximal inspiration; a “blast” of exhalation; and continued complete exhalation to the end of test (Miller et al., 2005). On spirometers a plot called a spirogram is generated at the end of each manoeuvre, measuring air flow. This is typically presented to health professionals as a volume-time (Figure 1) and a flow-volume graph. A FEM requires the coaching of the patient by a specialized health professional, due to the quality and repeatability criteria that must be met (Miller et al., 2005). Assuring that these criteria are fulfilled is of paramount importance, as neglecting them has led to over 25% of false-positives in diagnosing chronic obstructive pulmonary disease

(Moger et al., 2013), and 50% of false-negatives (Walters et al., 2011).




Figure 1: Volume-time curve showing the calculation of the Back Extrapolated Volume.

One of those criteria is the satisfactory start of exhalation, measured by the back extrapolated volume being higher than 5% of the Forced Vital Capacity (the total airflow in litres exhaled by the patient) or 150mL, whichever is greater (Miller et al., 2005). The back extrapolated volume is the exhaled volume at the instant where the maximum derivative of the volume-time curve crosses the abscissa axis (Figure 1).

Due to the growing popularity of smart devices, work has been developed to enable accurate estimation of FEM medical parameters outside of a clinical setting (or in otherwise resource constrained settings) making use of their computational capabilities and embedded sensors, especially the microphone (Larson et al., 2012), (Stein, 2013), (Liu, 2013), (Teixeira et al., 2015), (Zubaydi, 2016). However, there lies a largely unexplored problem common to all these solutions, which is that of assuring the validity of the manoeuvre in the absence of a health professional. In 3 of these works, we have

Improvements in automatic spirometry quality analysis have been recently developed for clinical spirometers (Melia et al., 2014), (Luo et al., 2017), but they do not take into account specific challenges faced by FEM acquired by microphone, henceforth referred to as microphone spirometry. To the best of our knowledge, no work has been done so far in the field of automatic quality evaluation in microphone spirometry.

This paper presents a first attempt of automatic quality evaluation in microphone spirometry, specifically on an initial effort criterion based on the ATS/ERS quality criteria (Miller et al., 2005). Adequate initial effort is already difficult even with the presence of a specialized health professional incentivizing the child. In the absence of such personnel, it becomes even more critical to correctly determine if the patient exhaled with enough force.

The development of this automatic quality evaluation module is part of a mobile serious game called "Ar.cade". It is a virtual pet game, with an asthmatic dragon. Its purpose is to allow and incentivize long term asthma monitoring in children from 5 to 10 years old, away from their healthcare professional, via microphone spirometry with smart mobile devices. While more typical actions such as feeding and cleaning a virtual pet will be available to the player, the main focus is on the mini-games. These revolve around the usage of the microphone as the main game input, rewarding the player for properly executed FEM and providing feedback on how to improve, in case of failed quality criteria.

2 EXPERIMENTAL SETUP

The assessment of identification methods requires a properly annotated database of microphone spirometry recordings. Therefore, one mini-game of Ar.cade was used for data collection and classification.

2.1 Game Design

Ar.cade is an Android mobile virtual pet game, developed and implemented in C# using the FlatRedBall game engine. Among other things, inside it can be found mini-games that use the microphone as the main controller, for the purpose of recording FEM's. The selected mini-game is a physics-based game, using the Farseer physics engine. The player character is a dragon, which is able to make a fireball-like projectile with its breath (Figure 2).




Figure 2: The game's idle state.

The main game loop consists of a 5 main phases:

- 1 Inhale phase (Figure 3): the screen zooms in on the dragon, a countdown with visible and audible feedback starts, and at the same time the dragon performs an animation to inhale deeply. As the countdown approaches the end, other background sound effects are gradually muted.




Figure 3: The game's inhale phase.

- 2 Exhale phase (Figure 4): Having all existing game sounds muted and the countdown finished, the

audio recording starts. The dragon performs an exhalation animation for a total of 3.5 seconds, after which the recording stops.




Figure 4: The game's exhale phase.

During this animation, fire particles are blown by the dragon and a slowly expanding projectile starts to form.

- 2 Upgrade phase (Figure 5): Quality of the manoeuvre is reflected by upgrading or downgrading projectile. For now, the final state is determined by the type of test the healthcare professional intends to perform, but in the future the quality detection module will evaluate this automatically.




Figure 5: The game's upgrade phase.

- 3 Destruction phase (Figure 6): The projectile is launched towards a destructible structure built of blocks with different materials.



Figure 6: The game's destruction phase.

Feedback phase (Figure 7): The player receives an award based on his/her performance. In the future, this will be linked with the overarching virtual pet game's economy, however currently just serves as another potential instant gratification source. More importantly, this will be the place where the player will receive instructions on how to improve their manoeuvre and receive the next possible ranking, in case of a sub-optimal manoeuvre.




Figure 7: The game's feedback phase.

The dragon's inhale and exhale animations serve the purpose of incentivizing the child to perform maximal effort on both phases, while the rest of the gameplay elements are an attempt to provide the child with instant gratification for the effort made.

2.2 Audio Processing Pipeline

To extrapolate the flow-time chart from an audio capture, the processing pipeline (Figure 8) as presented in (Teixeira et al., 2015) was implemented in C# for integration with the Ar. code project. In that work, an attempt was made to measure and classify lung function based on signal processing, constructing the flow-time curve. This would then be followed by a machine learning stage that enabled the regression of typical spirometry parameters. To perform this regression, a previously obtained database from adults was used for model training purposes. Given that we have no identical database for users in our target age group, we were unable to attempt a similar approach – that is, to try and establish absolute medical parameters. We have chosen to rely only on relative spirometry criteria for this work.

Automatic signal segmentation precedes the pre-processing stage. This serves to remove non-expiration sounds from the input to be analysed: a modified version of the back-extrapolation algorithm was used to determine the initial instant, and a sliding window algorithm based on the magnitude ratio threshold to determine the end (Teixeira et al., 2015).

The pre-processing stage attempts to transform the raw pressure data obtained from the microphone into the airflow measured at the lips. Afterwards, the envelope of the signal is extracted, with smoothing being applied on the post-processing stage.

3 DATA COLLECTION

3.1 Participants

The target population were children with/without asthma, aged between 5 to 10 years old. Data collection occurred in an informal environment, namely with the collaboration of a school.

Data was gathered anonymously, with written permission of legal guardians of all the children.

Besides the FEM audio recordings, self-reported data comprised of sex, ethnicity, age, and if they had asthma. Each child was assigned an internal random ID, enabling anonymous same-child recording analysis and comparison.

3.2 Procedures

A specialized healthcare professional performed the data collection. After introducing the child to the game's concept and performing a demonstration of the game, two different tests were made:

- "Hot air" test: Have the child exhale with a wide open mouth, focusing on achieving a good aperture and not emphasizing the need to exhale with maximal force.
- Maximal force test: Have the child exhale with the same mouth aperture, only this time with the added requirement and emphasis of maximal force, evaluated by the healthcare professional.

For each of these tests, the goal was to achieve at least one successful recording. At the end of each maneuver, the healthcare professional registered its quality with an in-game form assessing the maneuver on 6 different criteria in a yes/not sure/no format:

- Good mouth aperture
- Good initial effort
- Good continuous effort
- Good finish
- No cough/outside interference
- No glottis closure

3.3 Algorithm Development and Evaluation

Official guidelines defined by the ATS (Miller et al.,

2005) state that BEV should be lower than 150mL or 5% of the FVC, whichever is higher. As a first attempt to automatically determine if the manoeuvre's initial effort was acceptable or not, the official guidelines were used in as much as possible, and we implemented an algorithm for BEV calculation according to ATS standards (Miller et al., 2005).

Given that we do not have access to absolute values in our implementation of the processing pipeline, we chose to only use the relative criterion of $BEV < 5\%$ of FVC. We then compared the results obtained by this classifier with the healthcare professional's classification of the manoeuvres, acting as our ground truth.




Figure 8: In orange, the acceptable BEV ranges using the ATS guidelines; in blue, the ranges using the implemented algorithm (The ATS ranges overlap with the algorithm's ranges).

In figure 8 is shown that for FVC values under 3 litres, the ATS guidelines are increasingly more lenient as the FVC decreases compared to just using the relative criteria implemented in the algorithm. To evaluate how relevant this issue is for our collected data, we used children's age specific reference equations for FVC developed by (Koopman et al., 2011). According to these equations, FVC varies with age, height and sex. We used the self-reported age and sex, while for height the World Health Organization's height-for-age charts were considered (de Onis et al., 2007) ("WHO | Height-for-age (5-19 years)," n.d.).

To illustrate the FVC boundaries of our targeted population the mean FVC, along with the lower and upper limit of normal (LLN and ULN) for a 5% cut-off, are shown in tables 1 and 2.

Table 1: FVC percentiles (5% cut-off) for 10-year-old children in the 99th height percentile.

	LLN	Mean	ULN
Male	2.52L	3.02L	3.65L
Female	2.35L	2.86L	3.41L

This allows us to conclude that, for our target age group, it is more likely to be dealing with cases where

the expected FVC is under 3L, making our algorithm more stringent than the official ATS guidelines.

Table 2: FVC percentiles (5% cut-off) for 5-year-old children in the 1st height percentile.

	LLN	Mean	ULN
Male	0.73L	0.91L	1.11L
Female	0.82L	1.00L	1.19L

As an initial attempt to overcome this limitation, we defined an age adjusted BEV threshold for each of the target ages. This threshold was calculated as thus:

$$BEV_i = 0.15/FVC_i \quad (1)$$

where FVC_i is the mean expected FVC value for the given children's age (i), assuming the 50th height percentile of the WHO standards. The resulting BEV thresholds for males and females can be seen on table 3 and 4, respectively. By doing this, we are trying to have the acceptable BEV for a specific age match the static 150mL criterion defined by the ATS, but by using the FVC percentage relative criterion.

The value found for 5-year-old males is in complete agreement with the findings of (Aurora 2004) in pre-schoolers observed BEV/FVC ratio. In that work, a possible quality control cut-off of 12.5% is suggested.

Table 3: Age adjusted BEV thresholds for the 50th height percentile of each presented age in males.

Age	5	6	7	8	9	10
BEV (%)	12.5	10.7	9.3	8.2	7.0	6.4

Table 4: Age adjusted BEV thresholds for the 50th height percentile of each presented age in females.

Age	5	6	7	8	9	10
BEV (%)	12.9	11.1	9.8	8.6	7.5	6.6

Considering that the target is the classification made by the healthcare professional, where positives indicate an acceptable manoeuvre, the prediction is the resultant classification by the algorithm for the specified BEV threshold can be evaluated.

To compare different settings for the algorithm, two measures were used, namely the F_1 score and accuracy. The F_1 score is the harmonic mean of precision and sensitivity given by:

$$F_1 = 2TP / (2TP + FP + FN) \quad (2)$$

and the accuracy, which measures how often is the classifier correct, with the following formula:

$$A = (TP + TN) / (TP + TN + FP + FN) \quad (3)$$

4 RESULTS AND DISCUSSION

4.1 Participants

A total of 55 children between 5 to 10 years old have participated, 52 within 8 to 10 years of age. Out of these children, 31 were females and 24 were males. Only 4 (7.2%) of these children reported to suffer from asthma.

In Table 5 we present the distribution of the classifications by the healthcare professional, along with the total amount of recordings obtained for each type of test. In this case, "yes" represents a positive evaluation of the child's initial effort. Table 6 characterizes the amount of manoeuvres required by the healthcare professional to obtain a positive classification.

Table 5: Total amount of recordings obtained per test, and the distribution of acceptable BEV quality classification by the healthcare professional registered quality.

Test Type	Total	Yes	Not Sure	No
Hot Air Test	89	65	2	22
Max Force Test	144	108	3	33

Table 6: Number of attempts until a positive classification by the healthcare professional.

Test type	Max	Median	Min
Hot Air Test	3	1	1
Max Force Test	5	1	1

4.2 Quality Assessment

Tables 7 and 8 show the confusion matrix for the algorithm using ATS guidelines relative BEV threshold and the healthcare professional, for the two tests performed.

Table 7: Confusion matrix for the hot air tests using the BEV threshold <5% FVC.

Hot Air Test		Predicted	
BEV < 5%		Yes	No
Target	Yes	27.3%	47.6%
	No	7.1%	17.8%

Table 8: Confusion matrix for the maximal effort tests using the BEV threshold <5% FVC.

Max Force Test		Predicted	
BEV < 5%		Yes	No
Target	Yes	36.2%	41.3%
	No	18.1%	13.0%

At a first glance, the high false negative rates indicate that the algorithm is too strict with respect to the healthcare professional. Nevertheless, given that the only information available to the healthcare professional to produce an evaluation was the visual observation of the child, and not an objective measurement obtained by a spirometer as would happen on a regular spirometry, it is certainly possible that there was some mislabelling – specifically by accepting manoeuvres that otherwise would not have been accepted.

Given the already explained influence of age, sex and height on the FVC, it is worth noting that the presented results are based off a sample database that is heavily biased towards the higher end of the age spectrum. Given the positive correlation between age and height to expected FVC values, this means that it would be reasonable to expect our results to be worse with a more balanced database, for any static BEV threshold.

Table 9: Confusion matrix for the hot air tests using the BEV threshold <12.5% FVC.

Hot Air Test BEV < 12.5%		Predicted	
		Yes	No
Target	Yes	64.2%	10.7%
	No	22.6%	2.4%

Table 10: Confusion matrix for the maximal effort tests using the BEV threshold <5% FVC.

Max Force Test BEV < 12.5%		Predicted	
		Yes	No
Target	Yes	68.8%	8.7%
	No	18.1%	4.3%

As a test, we repeated the analysis for the BEV threshold suggested by (Aurora et al., 2004), that coincided with the estimated male BEV threshold for 5-year olds on Table 3, 12.5% (Tables 9 and 10). While it did improve the true positives and false negatives, it came at a cost of true negatives, and false positives in the case of the hot air test. However, it is important to note that the hot air test is of lesser importance compared to the maximal effort test, not only for the scope of this paper, but for Ar.cade's scope: the players are expected to have been coached to perform maximal effort manoeuvres, whereas they were not in the hot air test. This is due to specific instructions and coaching for maximal force on exhale being only given on the maximal effort test, and not before. In terms of gameplay, a false negative would mean asking the child to repeat the manoeuvre, while a false positive would promote poor form. Taken to the extreme, both of these would lead to the

failure of the "Ar.cade" project, even if due to different reasons: high difficulty causing frustration and a loss of interest in the game, or useless gathered results from a medical standpoint. Therefore, a hybrid approach might be worth exploring: starting out with a lower BEV threshold, but raising the threshold after several failed attempts.

The Tables 11 and 12 show the confusion matrixes for the age adjusted BEV thresholds. When compared to the results of the static BEV thresholds, they are somewhere in between them. The differences between these are cleared when looking at Tables 13 and 14, which present the F1 score and accuracy measurements for the 2 presented static BEV thresholds along with the age adjusted approach. As expected, the 5% threshold gives the worst results. The 12.5% threshold appears to be best, and while previously mentioned research does point to this threshold as appropriate for the younger children in our target group, it is important to note this: F1 score and accuracy can provide falsely inflated results in unbalanced classes, such as the ones presented in our confusion matrices.

Table 11: Confusion matrix for the hot air tests using the age adjusted BEV thresholds.

Hot Air Test Age adjusted BEV		Predicted	
		Yes	No
Target	Yes	50.0%	25.0%
	No	15.5%	9.5%

Table 12: Confusion matrix for the maximal force tests using the age adjusted BEV thresholds.

Max Force Test Age adjusted BEV		Predicted	
		Yes	No
Target	Yes	55.1%	22.4%
	No	14.5%	8.0%

Table 13: F1 scores, accuracy measurements for different BEV thresholds in the hot air test.

BEV	F1 score	Accuracy
5%	50.0%	45.2%
12.5%	79.4%	66.6%
Age adjusted	71.1%	59.5%

Table 14: F1 scores, accuracy measurements for different BEV thresholds in the maximal force test.

BEV	F1 score	Accuracy
5%	58.8%	49.2%
12.5%	83.7%	73.2%
Age adjusted	74.9%	63.0%

Aside from this, according to (Koopman et al.,

2011) the direct correlation between FVC and age is small compared to FVC's correlation with height, as can be seen on Figure 9. Given this notable discrepancy, and the fact that we assumed the average height for each age shows another significant limitation in our work, as the height for a child varies around 20% from the 3rd percentile to the 97th, at any given age in our target group according to WHO standards.




Figure 9: Plot illustrating the importance of the predictors for FVC estimation as they vary with age or height. The red and blue lines represent the impact of children's height, measured in cm, for females and males respectively (target group bounded between 100 and 150cm, using WHO's height-to-age standards). The green line represents the impact of age, measured in months, for both females and males (target group between 60 and 120 months of age).

Table 15: Number of children capable of performing at least one maximal force test with acceptable BEV for different thresholds, in absolute and relative units respectively.

Criterion	Children with at least one positive classification
Healthcare professional	54 (98%)
5% FVC	46 (84%)
12.5% FVC	55 (100%)
Age adjusted	52 (95%)

Despite the limitations described above, it is important to note. On Table 15, we can see the number of children that managed at least one positive detection with the current algorithm, set at different thresholds, considering the maximal force tests which are more relevant to us as already explained. Even for our worst performing criterion, we obtained reasonably similar results to (Tomalak et al., 2008), where 80.4% of 117 children between the ages of 4 and 10 years old were able to pass the ATS standard for BEV acceptability, using clinical spirometers.

This shows that while much work is still needed, for a first approach on the unexplored field of microphone spirometry automatic quality assessment (at least to the best of our knowledge), the results look promising.

4.3 Limitations and Future Work

Given the source of children for this study, only 7.2% of them had asthma. It is important to have a larger representation of these cases to establish how much can we extrapolate from studies in healthy children, and what are the specific challenges present in asthmatic children.

The fact that our ground truth was established by a single healthcare professional may have introduced a bias in our database, and to reduce this risk further data collection events should be performed with different healthcare professionals, with them cross-evaluating the same manoeuvre.

As was already mentioned, the database was heavily biased in terms of age distribution. Therefore, in further data collection events, there should be an increased focus on gathering audio samples from children under 7 years of age. This is especially important to further test the validity of the age adjusted BEV approach, to evaluate whether it has any merit to it.

5 CONCLUSIONS

The automatic evaluation of the FEM through BEV estimated from microphone spirometry allowed the assessment of the manoeuvre's quality, with respect to the start of exhalation.

Using the alternative less strict BEV thresholds of 12.5% and the Age Adjusted version, the quality was correctly assessed for over 70% of the manoeuvres. At least one acceptable manoeuvre was achieved for 96% of the children. Even using the stricter criteria of 5% of FVC it was possible to ensure at least one acceptable manoeuvre for 69%, which is slightly lower with the reported spirometry quality ratio in the literature for this age group.

While this leads us to conclude that our results are acceptable, at the same time we recognize the need of improvements for an automated system like this to become feasible in a real-world application.

ACKNOWLEDGEMENTS

We would like to thank the Externato das Escravas do Sagrado Coração de Jesus for their hospitality and cooperation, along with the parents of the 42 children for allowing them to participate in this study, enabling us to further the work in microphone spirometry.

The authors of this paper thank Susana Fonseca, R&D director of MEDIDA, for her continued support.

C. Jácome and R. Amaral have a post-doctoral (SFRH/BPD/115169/2016) and a PhD grant (PD/BD/113659/2015), respectively, funded by Fundação para a Ciência e a Tecnologia (FCT), Portugal.

R. Almeida, R. Guedes and J. P. Teixeira are part of NANOSTima - Project NORTE-01-0145-FEDER-000016 (NanoSTIMA), financed by the NorthPortugal Regional Operational Programme (NORTE 2020), under the PORTUGAL 2020 Partnership Agreement, and through the European Regional Development Fund (ERDF).

REFERENCES

- Aurora, P., Stocks, J., Oliver, C., Saunders, C., Castle, R., Chaziparasidis, G., Bush, A., 2004. Quality Control for Spirometry in Preschool Children with and without Lung Disease. *Am. J. Respir. Crit. Care Med.* 169, 1152–1159. <https://doi.org/10.1164/rccm.200310-1453OC>
- C. Larson, E., Goel, M., Boriello, G., Heltshe, S., Rosenfeld, M., N. Patel, S., 2012. SpiroSmart: Using a microphone to measure lung function on a mobile phone. *UbiComp12 - Proc. 2012 ACM Conf. Ubiquitous Comput.* <https://doi.org/10.1145/2370216.2370261>
- de Onis, M., Onyango, A.W., Borghi, E., Siyam, A., Nishida, C., Siekmann, J., 2007. Development of a WHO growth reference for school-aged children and adolescents. *Bull. World Health Organ.* 85, 660–667.
- Koopman, M., Zanen, P., Kruitwagen, C.L.J.J., van der Ent, C.K., Arets, H.G.M., 2011. Reference values for paediatric pulmonary function testing: The Utrecht dataset. *Respir. Med.* 105, 15–23. <https://doi.org/10.1016/j.rmed.2010.07.020>
- Liu, X., 2013. mCOPD: Mobile Phone Based Lung Function Diagnosis and Exercise System for COPD. *UCLA*.
- Luo, A.Z., Whitmire, E., Stout, J.W., Martenson, D., Patel, S., 2017. Automatic characterization of user errors in spirometry, in: 2017 39th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC). Presented at the 2017 39th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC), pp. 4239–4242. <https://doi.org/10.1109/EMBC.2017.8037792>
- Melia, U., Burgos, F., Vallverdú, M., Velickovski, F., Lluch-Ariet, M., Roca, J., Caminal, P., 2014. Algorithm for Automatic Forced Spirometry Quality Assessment: Technological Developments. *PLOS ONE* 9, e116238. <https://doi.org/10.1371/journal.pone.0116238>
- Miller, M.R., Hankinson, J., Brusasco, V., Burgos, F., Casaburi, R., Coates, A., Crapo, R., Enright, P., Grinten, C.P.M. van der, Gustafsson, P., Jensen, R., Johnson, D.C., MacIntyre, N., McKay, R., Navajas, D., Pedersen, O.F., Pellegrino, R., Viegi, G., Wanger, J., 2005. Standardisation of spirometry. *Eur. Respir. J.* 26, 319–338. <https://doi.org/10.1183/09031936.05.00034805>
- Moger, A., Holton, K., Hill, S., Kearney, M., Winter, R., 2013. Quality Assured Diagnostic Spirometry(QADS) - Performance And Competence, in: B45. CHRONIC OBSTRUCTIVE PULMONARY DISEASE: DIAGNOSIS AND EVALUATION, American Thoracic Society International Conference Abstracts. American Thoracic Society, pp. A2836–A2836. https://doi.org/10.1164/ajrcm-conference.2013.187.1_MeetingAbstracts.A2836
- Pierce, R., 2005. Spirometry: an essential clinical measurement. *Aust. Fam. Physician* 34, 535–539.
- Stein, B. van, 2013. A Mobile Smart Care platform Home spirometry by using the smartphone microphone.
- Teixeira, J.F., Teixeira, L.F., Fonseca, J., Jacinto, T., 2015. Automatic Analysis of Lung Function Based on Smartphone Recordings, in: Biomedical Engineering Systems and Technologies, Communications in Computer and Information Science. Presented at the International Joint Conference on Biomedical Engineering Systems and Technologies, Springer, Cham, pp. 390–402. https://doi.org/10.1007/978-3-319-27707-3_24
- Tomalak, W., Radliński, J., Latawiec, W., 2008. Quality of spirometric measurements in children younger than 10 years of age in the light of the recommendations. *Pneumonol. Alergol. Pol.* 76, 421–425.
- Walters, J.A., Walters, E.H., Nelson, M., Robinson, A., Scott, J., Turner, P., Wood-Baker, R., 2011. Factors associated with misdiagnosis of COPD in primary care. *Prim. Care Respir. J. J. Gen. Pract. Airw. Group* 20, 396–402. <https://doi.org/10.4104/pcrj.2011.00039>
- WHO | Height-for-age (5-19 years) [WWW Document], n.d.. WHO. URL http://www.who.int/growthref/who2007_height_for_age/en/ (accessed 4.3.18).
- Zubaydi, F.K., 2016. A mobile Based Platform for Monitoring Respiratory Diseases (Thesis).